

Certificate of Analysis

Product Name:Staphaurex KT/400 TSTProduct Code:R30859902Lot Number:6246675Expiration Date:20.11.2026Date of Manufacture:30.06.2025

*Date format is DD.MM.YYYY

Item W74TZL30M	Description Staphaurex Latex Filled Bulk	Lot/Serial 6234344	Expire 20.11.2026	
Test results carried out according to the Quality Assurance Product Release and Test Specification				
Test	Anticipated Result	Actual Result		
Smoothness	Test latex must not show agglutination although a trace granularity is acceptable	Conforms		
	Test Latex should spread easily to cover the circle	Conforms		
Efficacy/ Specificity	Test Latex must show obvious agglutination with the appropriate positive Staphylococcus cultures	Confo	Conforms	
	Test Latex must not show agglutination with the appropriate negative Staphylococcus cultures (trace granularity is acceptable)	Confo	rms	

W.

Fay Curd Quality Assurance Manager