

Certificate of Analysis

Product Name: Staphaurex Plus KT/450TST
Product Code: R30950201
Lot Number: 6227227
Expiration Date: 04.04.2026
Date of Manufacture: 25.04.2025

*Date format is DD.MM.YYYY

| Item | Description | Lot/Serial | Expire |
|-----------|--------------------------------------|------------|------------|
| W74CZL33E | Staph Plus Control Latex Filled Bulk | 6196412 | 04.04.2026 |
| W75DZL33F | Staph Plus Test Latex Filled Bulk | 6196413 | 04.04.2026 |

Test results carried out according to the Quality Assurance Product Release and Test Specification

| Test | Anticipated Result | Actual Result |
|--------------------------|--------------------------------------------------------------------------------------------------------------------------------------------|---------------|
| Smoothness | Test and Control latex must not show agglutination although a trace granularity is acceptable | Conforms |
| | Test and Control Latex should spread easily to cover the circle | Conforms |
| Efficacy/ Specificity | Test Latex must show obvious agglutination with the appropriate positive Staphylococcus cultures | Conforms |
| | Control Latex must not show agglutination | Conforms |
| | Test and Control Latex must not show agglutination with the appropriate negative Staphylococcus cultures (trace granularity is acceptable) | Conforms |



Fay Curd
Quality Assurance Manager