

Certificate of Analysis

Product Name: Streptex Latex Gp B 50TST/KT
Product Code: R30950701
Lot Number: 6229757
Expiration Date: 31.05.2026
Date of Manufacture: 16.04.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZL52W	ZL52 Strep B Latex Filled Bulk	6190214	31.05.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Abbreviated Latex Efficacy/Specificity	The Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute extraction procedures	Conforms
	The Latex must not show agglutination with the heterologous extracts although a trace granularity is acceptable	Conforms
Compatability of Latex with Streptex Acid Extraction kit	The latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts nor with the extraction fluid from the uninoculated tube although a trace granularity is acceptable	Conforms



Fay Curd
Quality Assurance Manager