

Certificate of Analysis

Product Name: Streptex Acid Extraction KT/50TST
Product Code: R30951301
Lot Number: 6227226
Expiration Date: 18.02.2027
Date of Manufacture: 11.06.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74AZL59A	Extraction Reagent 1	6192754	18.02.2027
W74AZL59B	Extraction Rgt 2	6192755	20.02.2027
W74AZL59C	Extraction Reagent 3	6192756	19.02.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Efficacy of Extraction Reagents 1, 2 and 3	Each Latex must show obvious agglutination with its homologous extracts	Conforms
	Each Latex must not show agglutination with the heterologous extracts, nor with the extraction fluid from the uninoculated tube although trace granularity is acceptable	Conforms



Fay Curd
Quality Assurance Manager