

Certificate of Analysis

Product Name: Salmonella O Factor 4 (Gp B) 2ml/VL
Product Code: R30956901
Lot Number: 6228204
Expiration Date: 11.12.2027
Date of Manufacture: 19.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZC13P	Salmonella 4-O Antisera Filled Bulk	6179299	11.12.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Identity Slide Agglutination	Saline control should give a negative reaction	Conforms
	Heterologous strain must give a negative reaction (trace is acceptable)	Conforms
	Homologous strain must give a positive reaction equal to or greater than +	Conforms



Fay Curd
 Quality Assurance Manager