

## Certificate of Analysis

**Product Name:** Salmonella O Factor 9 (Gp D) 2ml/VL  
**Product Code:** R30957301  
**Lot Number:** 6253410  
**Expiration Date:** 02.06.2028  
**Date of Manufacture:** 26.06.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZC17W	Salmonella 9-O Antisera Filled Bulk	6240993	02.06.2028

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Abbreviated Identity Slide Agglutination	Saline control should give a negative reaction	Conforms
	Heterologous strain must give a negative reaction (trace is acceptable)	Conforms
	Homologous strain must give a positive reaction equal to or greater than +	Conforms



Fay Curd  
 Quality Assurance Manager