

Certificate of Quality

Product Name: OADC Enrichment (100ml) each
Lot Number: 281948

Product Number: R450603
Expiration Date: 2026-06-11
(YYYY-MM-DD)


This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Mycobacterium tuberculosis	25177	Good Growth
Mycobacterium kansasii	12478	Good Growth
Mycobacterium scrofulaceum	19981	Good Growth
Mycobacterium intracellulare	13950	Good Growth
Mycobacterium fortuitum	6841	Good Growth

The enrichment is added to 7H10 Deep 09565 for Quality Control testing to be performed.

Physical Characteristics:
Appearance: Light to Medium Amber, Clear Broth
pH @ 20-25°C: N/A

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed


Senior Quality Assurance Engineer