

Certificate of Quality

**Product Name:** Sab Dex Agar,pH 5.6-500g each  
**Lot Number:** 248474

**Product Number:** R454462  
**Expiration Date:** 2029-04-12  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.


Organism Challenge	ATCC®	Performance Criteria
Aspergillus brasiliensis	16404	Good Growth
Candida albicans	10231	Good Growth
Cryptococcus neoformans	34877	Good Growth
Trichophyton mentagrophytes	9533	Good Growth

Visual Driform: Light beige, homogenous, free-flowing powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 5.6 +/- 0.2

This product meets or exceeds USP and CLSI growth promotion requirements.

This lot meets criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed  


Senior Quality Assurance Engineer