

## Certificate of Analysis

**Product Name:** PathoDx Herpes Typing KT/100TST  
**Product Code:** R62250  
**Lot Number:** 6183543  
**Expiration Date:** 31.12.2025  
**Date of Manufacture:** 16.04.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74IFA01A	Mounting Fluid 5ml Filled Bulk	6127319	28.02.2026
W74IFA18C	HSV 1 Rgt 5.0ml	6148160	10.06.2026
W74IFA19D	HSV II Rgt 5.0ml	6148181	17.06.2026
W74IFA23J	Herpes Slides (pouched and labelled)	6183546	31.12.2025

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Required Result	Actual Result
Kit Compatibility Test	The HSV-I Typing Reagent must show fluorescence with the homologous HSV-1 well on the Herpes Typing Control Slide and show no fluorescence with the heterologous HSV-II well	Conforms
	The HSV-II Typing Reagent must show fluorescence with the homologous HSV-II well on the Herpes Typing Control Slide and show no fluorescence with the heterologous HSV-I well	Conforms



Fay Curd  
Quality Assurance Manager