

## Certificate of Quality

**Product Name:** RapID Inoculation Fluid - 1ml PK/20Tube  
**Lot Number:** 284158

**Product Number:** R8325102  
**Expiration Date:** 2027-06-15  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

This product has been tested for microbial content and aesthetic value only. Functional testing should be performed by the user to assure efficacy of the product.


**Physical Characteristics:**

Appearance:	Colorless, Clear Liquid
pH @ 20-25°C:	5.0 - 9.0

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer