

Certificate of Quality

Product Name: D/M WATER, 5ML, 100/BOX, IVD
Lot Number: 274278

Product Number: T3339
Expiration Date: 2027-05-21
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Performance Criteria:

Conductivity $\leq 50 \mu\text{S/cm}$


Physical Characteristics:

Appearance: Clear, Colorless Liquid

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer