

Certificate of Quality

**Product Name:** CAMHB W/TES, 11ML, 100/BOX, IVD  
**Lot Number:** 287627


**Product Number:** T3462  
**Expiration Date:** 2026-06-26  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	MIC limits are within CLSI specifications
Pseudomonas aeruginosa	27853	MIC limits are within CLSI specifications
Staphylococcus aureus	29213	MIC limits are within CLSI specifications
Enterococcus faecalis	29212	MIC limits are within CLSI specifications

**Physical Characteristics:**  
Appearance: Light Amber, Clear Broth w/ No Significant Precipitate  
pH @ 20-25°C: 7.3 -7.4

**Microbial Load Evaluation:**  
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed  


Senior Quality Assurance Engineer