

Certificate of Quality

Product Name: YEASTONE BROTH, 11ML, 10/BOX, IVD
Lot Number: 279317


Product Number: Y3462
Expiration Date: 2026-06-03
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|----------------------|-------|---|
| Candida parapsilosis | 22019 | MIC limits are within CLSI specifications |
| Candida krusei | 6258 | MIC limits are within CLSI specifications |

Physical Characteristics:
Appearance: Light Amber-Orange, Clear Broth w/ No Precipitate
pH @ 20-25°C: 7.0 +/- 0.1

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 7 days at 20-25C and 33-37C.

Signed


Senior Quality Assurance Engineer