

Certificate of Quality

Product Name: YEASTONE BROTH, 11ML, 10/BOX, IVD
Lot Number: 287010


Product Number: Y3462
Expiration Date: 2026-06-24
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Candida parapsilosis	22019	MIC limits are within CLSI specifications
Candida krusei	6258	MIC limits are within CLSI specifications

Physical Characteristics:	
Appearance:	Light Amber-Orange, Clear Broth w/ No Precipitate
pH @ 20-25°C:	7.0 +/- 0.1

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 7 days at 20-25C and 33-37C.

Signed


Senior Quality Assurance Engineer