

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Monterrey, NL, MEX.



Ruben Deschamps  
Sr. Quality Manager  
07/02/2025

The following information represents Product Certification for: Item#: **5115-0032**

Description: **LABTOP COOLER,-20DEG C,PC;32**

Lot#: **1421659**

Manufactured: **06/27/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-2530-24P	BASE,LG,MOLDED,PC BLUE.	COMPONENT PART				
8-0056-13P	RESIN,PC,BLU,INJ	COLOR MIX (PC, BLU, INJ)	N/A	N/A	N/A	N/A
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-09	COLOR,BLUE,PC,	COLORANT, BLUE	N/A	N/A	N/A	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-2530-91P	INSR,LG,MOLDED,PC BLUE.	COMPONENT PART				
8-0056-13P	RESIN,PC,BLU,INJ	COLOR MIX (PC, BLU, INJ)	N/A	N/A	N/A	N/A
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-09	COLOR,BLUE,PC,	COLORANT, BLUE	N/A	N/A	N/A	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-2530-23P	COV, OUTER,LG,LT COOLR,PC WHT	COMPONENT PART				
8-0056-12P	RESIN,PC,WHT,INJ	COLOR MIX (PC, WHT, INJ)	N/A	N/A	N/A	N/A
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-37	COLOR,WHITE,PC	COLORANT, WHITE	N/A	N/A	N/A	178.3297
1-2530-92P	COV, INNER,LG,LT COOLR,PC WHT	COMPONENT PART				
8-0056-12P	RESIN,PC,WHT,INJ	COLOR MIX (PC, WHT, INJ)	N/A	N/A	N/A	N/A
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-37	COLOR,WHITE,PC	COLORANT, WHITE	N/A	N/A	N/A	178.3297
1-2530-98P	INSR,0.5ML,MICROTUBE,LT COOLER	COMPONENT PART				
8-0070-50P	RESIN,PP,GRANITE,INJ	COMPONENT PART				

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Monterrey, NL, MEX.



Ruben Deschamps  
Sr. Quality Manager  
07/02/2025

The following information represents Product Certification for: Item#: **5115-0032**

Description: **LABTOP COOLER,-20DEG C,PC;32**

Lot#: **1421659**

Manufactured: **06/27/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
8-0071-06	Resin,PP,Inj	POLYPROPYLENE, INJECTION	9988	PASSED	PASSED	177.1520(a)(1)(i), (c)1.1a,177.1520(b), (use conditionsA-H)
8-0099-30	COLOR,GRANITE	COLORANT, GRANITE	N/A	N/A	N/A	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-2530-99	PLUG,LT COOLER,PC,WHT	COMPONENT PART				

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.