

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate
Thermo Scientific
Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Monterrey, NL, MEX.

Ruben Deschamps Sr. Quality Manager

06/17/2025

The following information represents Product Certification for: Item#: 5929-0030

Description: RACK TT PC;30MM Lot#: 1430913 Manufactured: 06/14/2025

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-2490-80	PLT,END,13/16/20/25/30MM,PC	COMPONENT PART				
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
1-2490-82	PLT,BTM,25/30MM,TT,RACK,PC,W HT	COMPONENT PART				
8-0056-12P	RESIN,PC,WHT,INJ	COLOR MIX (PC, WHT, INJ)	N/A	N/A	N/A	N/A
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-37	COLOR, WHITE, PC	COLORANT, WHITE	N/A	N/A	N/A	178.3297
1-2490-84	PLT,TOP/CNT,30MM,TT RACK,PC	COMPONENT PART				
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.