

ZERTIFIKAT CERTIFICATE

Pumpen | Probennehmer | Laborbedarf
für Labor, Industrie und Wissenschaft

Pumps | Sampling | Plastic Labware
for Laboratory, Industry, Science



Certificate of Compliance

7556-0025 Single-use pipette, PS crystal clear, sterile, 25ml
7556-0050 Single-use pipette, PS crystal clear, sterile, 50ml
7556-0100 Single-use pipette, PS crystal clear, sterile, 100ml

Lot ID: 28524056
Product reference: 7556-0100
Expiration Date: 2027-10-11
SAL Achieved: 10^{-3}

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-3} .

April 14, 2025

A handwritten signature in blue ink, appearing to read "M. Saint-Denis".

Bürkle GmbH, Bad Bellingen
Martin Saint-Denis, Managing Director