

ZERTIFIKAT CERTIFICATE

Pumpen | Probenehmer | Laborbedarf
für Labor, Industrie und Wissenschaft

Pumps | Sampling | Plastic Labware
for Laboratory, Industry, Science



Certificate of Compliance

7556-0025 Single-use pipette, PS crystal clear, sterile, 25ml

7556-0050 Single-use pipette, PS crystal clear, sterile, 50ml

7556-0100 Single-use pipette, PS crystal clear, sterile, 100ml

Lot ID:	35524025
Product reference:	7556-0050
Expiration Date:	2027-12-20
SAL Achieved:	10⁻³

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10⁻⁶.

April 30, 2025

A handwritten signature in blue ink, appearing to read "M. Saint-Denis".

Bürkle GmbH, Bad Bellingen
Martin Saint-Denis, Managing Director