

Corning Incorporated - Life Sciences  
2 Alfred Road  
Kennebunk ME 04043 USA  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: ERL, 5L, Plain, TRANSFER CAP, MPC, S, IND, 1/2		
Catalog Number	: 11501	Manufacture Date	: 2025-04-10
Lot ID	: 10025057		
Expiration Date	: 2028-04-09		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard and the FDA CFR 21 Part 820, current Good Manufacturing Practices (cGMP).

**BSE/TSE** - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Non-Pyrogenic** - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is  $\leq 0.25$  EU/ml or  $\leq 10$  EU/device.

**USP Class VI Testing** - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

**Sterility** - Products labeled Sterile Fluid Path have been designed to ensure sterility of the portion of the product intended for contact with fluids.

**Sterilization** - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-6}$ .

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass  
Packaging Inspection - Pass  
Integrity Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Renee E Gallagher  
Plant Quality Manager