

Certificate of Analysis

Catalog number	FB012927	Products name	6 well TC multidish	Inspection Date	2022/2/11
PO#	PR2589376	Origin	China	Lot Number	20211017-074-B
PO QTY (CS)	42	Subtotal (PCS)	4200	Expiry Date	2024-10-17

Inspection Item and Result							
Inspection Item and Standard		Acceptance Level			Inspection Result		
		Quality Level	Sample Size	Ac:Re	Defectiveness	Quantity	Judgement
Package	No (Incorrect/Illegible missing print, Incorrect/Missing label, Foreign matter/contam/insects, Under count (less bags in ctn), Damaged carton/boxes, Label and products can't match together.)	Level II; MA 0.65	8	0:1	0	0	ok
Visual (Cosmetics)	No (Scratch, Crack or deformation, Foreign Material, Dirt, Black spot, No transparency, Trimming)	Level II; MA 0.65	200	3:4	0	0	ok
Material	Virgin PS (FDA grade)	/	/	/	/	/	ok
Sterilization	SAL 10E-6, No bacteria growth	C=0	10	0:1	0	0	ok
Cell culture	Cell confluence >80%	C=0	3	0:1	0	0	ok
Cytotoxicity	Cell morphology, growth status is normal.	C=0	3	0:1	0	0	ok
Bacterial endotoxin	<0.5EU/ml	C=0	3	0:1	0	0	ok
Surface Treatment	Contact angle<45°	C=0	3	0:1	0	0	ok
Others	N/A						

The above merchandise is as per purchase order specifications

Final Judgement	☒	Yes	☐	No
	☒	Accept	☐	Reject

Signature

Inspected by (QA Inspector)	Approved by (QA Manager)	Reviewed by (SDE Manager)
Fiona Wang <i>Fiona Wang</i>	York Zhang <i>York Zhang</i>	Frem Wang <i>Frem Wang</i>

SHANGHAI, CHINA

Add: 2th Floor, Building 7, No.27, Xin Jinqiao Rd, Pudong (201206)

Tel: 021-61925688

Fax: 021-61002128

BEIJING, CHINA

Add: 7th Floor, Building F, Tower West, Yonghe Plaza, No.28, Andingmen East Street (100007)

Tel: 010-84193588

Fax: 010-84193580

GUANGZHOU, CHINA

Add: RM 204-206, North Tower, KWG The Star, No.36-38 Huanyu 3rd Road, International Bioisland (510000)

Tel: 020-82401600

Fax: 020-82401898