

Certificate of Analysis

Catalog number	FB012927	Products name	6 well TC multidish	Inspection Date	2023/9/6
PO#	PR3283395	Origin	China	Lot Number	20230804-074-F
PO QTY (CS)	10	Subtotal (PCS)	1000	Expiry Date	2026/08/04
Inspection Item and Standard					
Package	Incorrect/Illegible missing print				ok
	Incorrect/Missing label				ok
	Foreign matter/contam/insects				ok
	Under count (less bags in ctn)				ok
	Damaged carton/boxes				ok
	Label and products can't match together				ok
Visual (Cosmetics)	No Scratch, Crack or deformation				ok
	No Foreign Material or Black spot				ok
	No Dirt				ok
	Color: Transparency				ok
	No Trimming				ok
Material	Virgin PS (FDA grade)				ok
Sterilization	SAL 10 ⁻⁶ , No bacteria growth				ok
Cell culture	Cell confluence >80%				ok
Cytotoxicity	Cell morphology, growth status is normal.				ok
DNase & RNase	DNase & RNase Free				ok
Bacterial endotoxin	<0.5EU/ml				ok
Surface Treatment	Contact angle<45°				ok
Others	N/A				ok
The above merchandise is as per purchase order specifications					
Final Judgement		☒	Yes	☐	No
		☒	Accept	☐	Reject
Signature					
Inspected by (QA Inspector)		Approved by (Quality Director)		Reviewed by (SDE Manager)	
Fiona Wang <i>Fiona Wang</i>		Xiaoxiao Ma <i>Xiao Xiao Ma</i>		Frem Wang <i>long</i>	

SHANGHAI, CHINA

Add: 2th Floor, Building 7, No.27, Xin Jinqiao Rd, Pudong (201206)

Tel: 021-61925688

Fax: 021-61002128

BEIJING, CHINA

Add: 7th Floor, Building F, Tower West, Yonghe Plaza, No.28, Andingmen East Street (100007)

Tel: 010-84193588

Fax: 010-84193580

GUANGZHOU, CHINA

Add: RM 204-206, North Tower, KWG The Star, No.36-38 Huanyu 3rd Road, International Bioisland (510000)

Tel: 020-82401600

Fax: 020-82401898