

Certificate of Conformity

Sterile Product

Date: 17 Jun 2025

Product: 96 Flat Well Microtiter Plate

Item Number: 612F96

Lot Number: M566680

Date of Expiry: 2030/04

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.


Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10^{-6} .

	Min (kGy)	Max (kGy)	Certificate of Irradiation	Date of Irradiation
Calculated Dose	17.7	28.8	2163-12469A	14/15 Jun 25
Specification	15.0	35.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



Harvey Mayers

Interim Batch Release Document Controller