

## Certificate of Conformity

### Sterile Product

Date: 02 Jun 2025

**Product:** 500ml Water Sampling Cont. Undosed

**Item Number:** 500WSCNT

**Lot Number:** M563570

**Date of Expiry:** 2030/03

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

| Test Description | Requirements               | Result   |
|------------------|----------------------------|----------|
| Visual           | According to specification | Complies |
| Test of function | According to specification | Complies |

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL  $10^{-6}$ .

|                        | Min<br>(kGy) | Max<br>(kGy) | Certificate of<br>Irradiation | Date of<br>Irradiation |
|------------------------|--------------|--------------|-------------------------------|------------------------|
| <b>Calculated Dose</b> | 30.1         | 39.7         | 2163-12194A                   | 29/30 May 25           |
| <b>Specification</b>   | <b>27.5</b>  | <b>45.0</b>  |                               |                        |

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:

LIVIA HAVLICKOVA

Batch Release-Document Controller