

Certificate of Conformity

Sterile Product

Date: 02 Jun 2025

Product: 500ml Water Sampling Cont. Undosed

Item Number: 500WSCNT

Lot Number: M563590

Date of Expiry: 2030/03

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10^{-6} .

	Min (kGy)	Max (kGy)	Certificate of Irradiation	Date of Irradiation
Calculated Dose	30.1	39.7	2163-12194A	29/30 May 25
Specification	27.5	45.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:

LIVIA HAVLICKOVA

Batch Release-Documents Controller