

Certificate of Conformity**Non-Sterile Product**

Date: 02 Jun 2025

Product: 60mm Petri Dish Single Vent**Item Number:** 123-17**Lot Number:** M552540**Date of Expiry:** 2030/04

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001: 2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



LIVIA HAVLICKOVA

Batch Release-Document Controller