

Certificate of Conformity – Nalgene

Date: 25 Apr 2025

Product: Nalgene 125ml PET Square Bottle Capped
Item Number: 842040-0125
Lot Number: M559760
Resin Number: 8-0003-32
Closure Number: 430118-0320
Date of Expiry: 08 Apr 2030

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial Endotoxins Test	<0.5 endotoxins unit/ml	Complies
USP Class VI	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

	Min (kGy)	Max (kGy)	Certificate of Irradiation Ref	Date of Irradiation
Calculated Dose	22.6	33.8	45,283	23/24 Apr 25
Specification	20.0	45.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



LIVIA HAVLICKOVA
Batch Release-Document Controller