

Certificate of Conformity – Nalgene

Date: 11 Aug 2025

Product: Nalgene 650ml PET Square Bottle Capped
 Item Number: 842040-0650
 Lot Number: M571200
 Resin Number: 8-0003-32
 Closure Number: 430118-0320
 Date of Expiry: 29 Jul 2030

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.


Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial Endotoxins Test	<0.5 endotoxins unit/ml	Complies
USP Class VI	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

	Min (kGy)	Max (kGy)	Certificate of Irradiation Ref	Date of Irradiation
Calculated Dose	24.5	35.5	46,303 x 825	08 Aug 25
Specification	20.0	45.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



Harvey Mayers
Interim Batch Release-Document Controller