

Certificate of Conformity – Nalgene

Date: 30 May 2025

Product: Nalgene 1000ml PET Square Bottle Capped
 Item Number: 842040-1000
 Lot Number: M565310
 Resin Number: 8-0003-32
 Closure Number: 430118-0320
 Date of Expiry: 14 May 2030

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial Endotoxins Test	<0.5 endotoxins unit/ml	Complies
USP Class VI	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

	Min (kGy)	Max (kGy)	Certificate of Irradiation Ref	Date of Irradiation
Calculated Dose	24.8	37.5	45,624	29 May 25
Specification	20.0	45.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



LIVIA HAVLICKOVA
 Batch Release-Document Controller