

## Certificate of Conformity – Nalgene

Date: 02 Jun 2025

Product: Nalgene 1000ml PET Square Bottle Capped  
 Item Number: 842040-1000  
 Lot Number: M566350  
 Resin Number: 8-0003-32  
 Closure Number: 430118-0320  
 Date of Expiry: 14 May 2030

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.


Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial Endotoxins Test	<0.5 endotoxins unit/ml	Complies
USP Class VI	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10<sup>-6</sup>.

	Min (kGy)	Max (kGy)	Certificate of Irradiation Ref	Date of Irradiation
Calculated Dose	24.8	37.3	45,630 x 960	30 May 25
Specification	<b>20.0</b>	<b>45.0</b>		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



Harvey Mayers  
 Interim Batch Release-Documents Controller