

Certificate of Conformity

Date: 01 Jul 2025


Product: 30ml PP Container, DNase & HuDNA Free, Non-Sterile
Item Number: 30APPRN
Lot Number: M561730
Date of Expiry: 2030/06/11

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment (ISO Class 7), inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial endotoxins test	<0.01 endotoxins unit/ml	Complies
DNase Test	<1 x 10 ⁻⁴ U/ul	Complies
RNase Test	<5 x 10 ⁻⁸ U/ul	Complies
Human DNA Test	<5pg	Complies

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001: 2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:



Harvey Mayers

Interim Batch Release Document Controller