
Certificate of Conformity

Product: Nunc U96 MicroWell Plate, Natural**Date:** 2025-06-10**Page:** 1**Item number:** 267245**Lot number:** 186042**Date of manufacture:** April 2025**Date of expiry:** April 2030

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**
DNase/RNase test	Negative	Complies*

*Performed as monitoring test

**Only performed once

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature