

## CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

|                    |                                                |
|--------------------|------------------------------------------------|
| Short Description  | Non-sterile Disposable Nitrile Cleanroom Glove |
| Size(s)            | XL (9.5 - 10)                                  |
| Brand              | BioClean™                                      |
| Style              | Synergy BSAN                                   |
| Material           | Nitrile                                        |
| Lot Number         | 24-1913                                        |
| Batch Number       | N102408023                                     |
| Manufacturing Date | 08/2024                                        |
| Expiration Date    | 07/2029                                        |

### CERTIFICATE OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

| Standard/Normative Document        | Result   |
|------------------------------------|----------|
| Free from BSE/TSE (EMEA/410/01)    | Conforms |
| PPER (EU) 2016/425                 | Conforms |
| REACH REGULATION (EC) No 1907/2006 | Conforms |
| Sampling Plan ISO 2859-1           | Conforms |

### CERTIFICATE OF ANALYSIS

| Property                 | Test Method          | Specification     | Result   |
|--------------------------|----------------------|-------------------|----------|
| Water-Tight Test         | EN ISO 374-2         | G-1, AQL 1.5, PL2 | Conforms |
| Liquid Particle Count    | IEST-RP-CC005.4      | N=3, N/A          | Conforms |
| Extractable Ionic Burden | IEST-RP-CC005.4      | N=3, N/A          | Conforms |
| Dimension Test           | EN420                | S-2, 4.0          | Conforms |
| Physical Properties      | EN 455-2             | S-2, 4.0          | Conforms |
| Force at Break (N)       | EN 455-2             | ≥ 6               | Conforms |
| Packing Conformity check | Internal test method | N/A, N/A          | Conforms |
| Visual Inspection-major  | Internal test method | G-1, AQL 2.5      | Conforms |
| Visual Inspection-minor  | Internal test method | G1, AQL 4.0       | Conforms |

Electronically Signed by:

**Amy Jones**  
Manager, QA

Date: 23 September 2024

**UK, Europe, Middle East & Africa**

Nitritex Ltd.  
Ground Floor, 15 Kings Court  
Willie Snaith Road,  
Newmarket, Suffolk  
CB8 7SG,  
United Kingdom  
Tel: +44 (0) 1638 663338  
Email: nitritexuk@ansell.com

Ansell Healthcare Europe NV  
Riverside Business Park  
Bld International, 55  
1070 Brussels,  
Belgium  
Tel: +32 2 528 7400

**Asia Pacific**

Nitritex (M) Sdn Bhd  
No. 2, Jalan Jurunilai U1/20, Seksyen U1,  
Hicom Glenmarie Industrial Park,  
40150 Shah Alam, Selangor,  
Malaysia  
Tel: +603 5569 3857/3859  
Email: nitritexmalaysia@ansell.com

Ansell Limited  
678 Victoria Street  
Richmond, VIC, 3121  
Australia  
Tel: +61 3 9270 7270  
Email: anzcustomerservice@ansell.com

**North America**

Ansell Healthcare Products LLC  
111 Wood Avenue South, Suite 210  
Iselin, NJ 08830,  
United States  
US Tel: 1 800 800 0444  
CA Tel: 1 800 363 8340  
Email: customerserviceus@ansell.com

Ansell Commercial Mexico S.A. de C.V.  
Blvd. Bernardo Quintana 7001-C,  
Q7001 Torre II, Suites 1304-1306,  
Centro Sur, Querétaro, Qro. C.P. 76079,  
México  
Tel: +52 554 631 9776  
Email: cslac@ansell.com

Ansell Brazil Ltda.  
Rua das Figueiras 474,  
4º Andar - Bairro Jardim,  
Santo André - SP  
CEP: 09080-300  
Brasil  
Tel: +55-11-3356-3100  
Email: sac@ansell.com