

CERTIFICATE OF IRRADIATION

Brand	BioClean™
Size(s)	8.0
Material	Nitrile
Style	N-Plus BNPS
Short Description	Sterile Nitrile Cleanroom Glove
Lot Number	24-2026
Batch Number	N42408032
Manufacturing Date	08/2024
Expiration Date	07/2029

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	24-2026
Irradiation Date	10 September 2024
Min. Dose Delivered (kGy)	28.10
Max. Dose Delivered (kGy)	34.20

Electronically Signed by:

Derek Quek Yeang

Manager, QA

Date: 14 September 2024

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699