



## CERTIFICATE OF IRRADIATION

Brand	BioClean™
Size(s)	XL (9.5 - 10)
Material	Natural Rubber Latex
Style	Extra BLAS
Short Description	Sterile Latex Cleanroom Glove
Lot Number	24-3341
Batch Number	L102501001
Manufacturing Date	01/2025
Expiration Date	12/2029

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Synergy Sterilisation Rawang (M) Sdn. Bhd, Lot 42 Jalan Industrial 2/1, Rawang Integrated Industrial Park, Rawang, 48000 Malaysia
Irradiation Lot No	3043-2462A
Irradiation Date	9 January 2025
Min. Dose Delivered (kGy)	25.8
Max. Dose Delivered (kGy)	36.0

Electronically Signed by:  
**Derek Quek Yeang**  
Manager, QA  
Date: 17 January 2025

For additional information visit us at [www.ansell.com](http://www.ansell.com), or call us at:

### Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.  
T: +32 (0) 2 528 74 00  
F: +32 (0) 2 528 74 01

### Australia

Ansell Limited  
T: +61 1800 337 041  
F: +61 1800 803 578

### Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.  
T: +52 442 248 1544 / 248 3133

### North America Region

Ansell Healthcare products LLC  
T: +1 800 800 0444  
F: +1 800 800 0445

### Asia Pacific Region

Ansell Global Trading Center  
(Malaysia) Sdn Bhd  
T: +603 8310 6688  
F: +603 8310 6699