

CERTIFICATE OF IRRADIATION

| | |
|--------------------|-------------------------------------|
| Brand | BioClean-D™ |
| Size(s) | XS |
| Material | CleanTough™ |
| Style | S-BDCHT |
| Short Description | Sterile Drop-down garment with Hood |
| Lot Number | 25-0771 |
| Batch Number | 25-0771 |
| Manufacturing Date | 03/2025 |
| Expiration Date | 02/2028 |

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

| | |
|---------------------------------------|---|
| Min. Specified Irradiation Dose (kGy) | 25 |
| Irradiation Location | Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia |
| Irradiation Lot No | 25-0771 |
| Irradiation Date | 22 March 2025 |
| Min. Dose Delivered (kGy) | 27.10 |
| Max. Dose Delivered (kGy) | 35.80 |

Electronically Signed by:

Derek Quek Yeang

Manager, QA

Date: 27 March 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699