



CERTIFICATE OF IRRADIATION

Brand	BioClean™
Size(s)	8.5
Material	Nitrile
Style	Emerald BENS
Short Description	Sterile Nitrile Cleanroom Glove
Lot Number	23-2046
Batch Number	N62312027
Manufacturing Date	12/2023
Expiration Date	11/2028

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	23-2046
Irradiation Date	29 December 2023
Min. Dose Delivered (kGy)	26.90
Max. Dose Delivered (kGy)	35.20

Electronically Signed by:
Derek Quek Yeang
Manager, QA
Date: 5 January 2024

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Australia

Ansell Limited
T: +61 1800 337 041
F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.
T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC
T: +1 800 800 0444
F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center
(Malaysia) Sdn Bhd
T: +603 8310 6688
F: +603 8310 6699

Russia

Ansell PYC
T: +7 495 258 13 16