



CERTIFICATE OF IRRADIATION

Brand	BioClean-D™
Size(s)	3XL
Material	CleanTough™
Style	S-BDCCT
Short Description	Sterile cleanroom Coverall with Collar
Lot Number	25-1227
Batch Number	25-1227
Manufacturing Date	04/2025
Expiration Date	03/2028

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Synergy Sterilisation Rawang (M) Sdn. Bhd, Lot 42 Jalan Industrial 2/1, Rawang Integrated Industrial Park, Rawang, 48000 Malaysia
Irradiation Lot No	3043-3101A
Irradiation Date	2 May 2025
Min. Dose Delivered (kGy)	28.5
Max. Dose Delivered (kGy)	36.6

Electronically Signed by:
Derek Quek Yeang
Manager, QA
Date: 9 May 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Australia

Ansell Limited
T: +61 1800 337 041
F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.
T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC
T: +1 800 800 0444
F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center
(Malaysia) Sdn Bhd
T: +603 8310 6688
F: +603 8310 6699