



CERTIFICATE OF IRRADIATION

Brand	BioClean-D™
Size(s)	3XL
Material	CleanTough™
Style	S-BDSH
Short Description	Drop-down Sterile Coverall with Hood
Lot Number	25-0866
Batch Number	25-0866
Manufacturing Date	03/2025
Expiration Date	02/2028

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Synergy Sterilisation Rawang (M) Sdn. Bhd, Lot 42 Jalan Industrial 2/1, Rawang Integrated Industrial Park, Rawang, 48000 Malaysia
Irradiation Lot No	3043-2963A
Irradiation Date	9 April 2025
Min. Dose Delivered (kGy)	28.9
Max. Dose Delivered (kGy)	36.2

Electronically Signed by:

Derek Quek Yeang

Manager, QA

Date: 15 April 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699