

CERTIFICATE OF IRRADIATION

Brand	BioClean-D™
Size(s)	S
Material	CleanTough™
Style	S-BDFC
Short Description	Sterile cleanroom Coverall with Hood and integrated Boots
Lot Number	25-1423
Batch Number	25-1423
Manufacturing Date	05/2025
Expiration Date	04/2028

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	25-1423
Irradiation Date	19 May 2025
Min. Dose Delivered (kGy)	27.60
Max. Dose Delivered (kGy)	35.80

Electronically Signed by:

Derek Quek Yeang

Manager, QA

Date: 24 May 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699