

Certificate of Irradiation

MY01S12436493-1-1

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

NITRITEX (M) SDN BHD
No 2 Jalan Jurunilai U1/20
Seksyen U1 Hicom Glenmarie Ind Park
Selangor
40150 Shah Alam

Account Number:	101178								
Synergy Health Sales Part Reference:	1127486								
Customer Reference Number:	SNB20/0138								
Product Description:	ST Nit Sleeve 26 Port 6"(GSL15NITPP26)								
Validation Reference:	R160077 Rev.1								
Quantity Received:	1								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	40.0								
Other Process Details:	PACKING: 40Pieces/Ctn								
	<table border="1"> <thead> <tr> <th>PRODUCT CODE</th> <th>LOT NO</th> <th>BATCH NO</th> <th>QTY</th> </tr> </thead> <tbody> <tr> <td>GSL15NITPP26</td> <td>20-1163</td> <td>N362005020</td> <td>1</td> </tr> </tbody> </table>	PRODUCT CODE	LOT NO	BATCH NO	QTY	GSL15NITPP26	20-1163	N362005020	1
PRODUCT CODE	LOT NO	BATCH NO	QTY						
GSL15NITPP26	20-1163	N362005020	1						

Date and Time of Irradiation:	18-May-2020 07:44
Reference Dose Range kGy:	32.7 - 32.7
Calculated Minimum Dose kGy:	27.9
Calculated Maximum Dose kGy:	36.5

Irradiation Release Authorised By Synergy Sterilisation Rawang (M) Sdn. Bhd, a STERIS Company

Page 1 of 1