

<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 14-Nov-2019

MY01S12358952-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

NITRITEX (M) SDN BHD
No 2 Jalan Jurunilai U1/20
Seksyen U1 Hicom Glenmarie Ind Park
Selangor
40150 Shah Alam

Order Information

Account Number:	101178			
Synergy Health Sales Part Reference:	1130611			
Customer Reference Number:	SNB19/0398			
Product Description:	St Nit Sleeve 26 Port 12" (GSL36NITPP26)			
Validation Reference:	R160077 Rev.1			
Quantity Received:	1			
Customer Minimum Specification kGy:	25.0			
Customer Maximum Specification kGy:	40.0			
Other Process Details:	PACKING : 40Pieces/Ctn			
	PRODUCT	LOT NO	BATCH NO	QTY
	CODE			
	GSL36NITPP26	19-3680	N411911013	1

Irradiation Data

Date and Time of Irradiation:	13-Nov-2019 19:08
Reference Dose Range kGy:	33.2 - 33.2
Calculated Minimum Dose kGy:	28.3
Calculated Maximum Dose kGy:	37.1

Irradiation Release Authorised By Synergy Sterilisation Rawang (M) Sdn. Bhd, a STERIS Company

Processing Site: Lot 42 Jalan Industrial 2/1, Rawang Integrated Industrial Park, Rawang, 48000 Phone No: +60(0)3 6099 9600

Registered Office: Suite 18.01, 18th Floor, MWE Plaza, No. 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number: 000878280704