

---

## Certificate of Conformity

**Product:** Nunc CryoFlex**Date:** 2025-06-11  
**Page:** 1**Item number:** 343958**Lot number:** 185962**Date of manufacture:** May 2025

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager  
Please note that this is an electronically generated signature