
Certificate of Conformity

Product: Nunc Immobilizer F96 MicroWell Plate,
Streptavidin, Clear

Date: 2025-07-03

Page: 1

Item number: 436014

Lot number: 186574

Date of manufacture: June 2025

Date of expiry: December 2026

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
Binding variation, tested by binding of biotinylated HRP	Plate CV% max 5 %	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature