

---

## Certificate of Conformity

**Product:** Nunc EasyFill Cell Factory Accessories, Filter Cap**Date:** 2025-06-17**Page:** 1**Item number:** 146008**Lot number:** 186569**Date of manufacture:** June 2025**Date of expiry:** June 2030

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

| Test Description  | Requirements               | Result     |
|---|----------------------------|------------|
| Visual  | According to specification | Complies   |
| Test of function  | According to specification | Complies   |
| USP <85> Bacterial endotoxins test                                  | <0.5 endotoxins units/mL   | Complies*  |
| USP <88> "Biological Reactivity test<br>In Vivo, USP Class VI Test" | USP current edition        | Complies** |

\*Performed as monitoring test

\*\*Only performed once

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature