
Certificate of Conformity

Product: Nunc Immobilizer F8 Module, Amino, Clear**Date:** 2025-06-18**Page:** 1**Item number:** 436013**Lot number:** 186525**Date of manufacture:** June 2025**Date of expiry:** June 2027

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

| Test Description | Requirements | Result |
|--|----------------------------|----------|
| Visual | According to specification | Complies |
| Test of function | According to specification | Complies |
| Binding variation, tested by coupling of C-Peptide | Plate CV% max 5% | Complies |

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature