
Certificate of Conformity

Product: Nunc Rack for CryoBank/Bank-it

Date: 2025-07-01

Page: 1

Item number: 367014

Lot number: 186743

Date of manufacture: June 2025

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature