

---

## Certificate of Conformity

**Product:** Nunc U96 MicroWell Plate, Natural**Date:** 2025-08-01**Page:** 1**Item number:** 267245**Lot number:** 186661**Date of manufacture:** July 2025**Date of expiry:** July 2030

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

| Test Description                                                    | Requirements               | Result     |
|---------------------------------------------------------------------|----------------------------|------------|
| Visual                                                              | According to specification | Complies   |
| Test of function                                                    | According to specification | Complies   |
| USP <88> "Biological Reactivity test<br>In Vivo, USP Class VI Test" | USP current edition        | Complies** |
| DNase/RNase test                                                    | Negative                   | Complies*  |

\*Performed as monitoring test

\*\*Only performed once

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature