
Certificate of Conformity

Product: Nunc 1 µl Inoculation Loop, PS, Clear**Date:** 2025-07-01**Page:** 1**Item number:** 253287**Lot number:** 186542**Date of manufacture:** June 2025**Date of expiry:** June 2030

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

The loops are manufactured to deliver a volume of 1 µL ± 0.3µL.

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature