

---

## Certificate of Conformity

**Product:** Nunc Universal Container, 25 ml w. writing area**Date:** 2025-06-18**Page:** 1**Item number:** 363282**Lot number:** 185655**Date of manufacture:** June 2025**Date of expiry:** June 2030

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
Leak Test	IATA current edition	Complies***

\*\*\*Performed at certain frequency

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature