

Certificate of Conformity

Product: Nunclon Sphera T25 Flask
Item Number: 174951
Lot number: 412502
Date of manufacture: March 2025
Date of expiry: March 2029

Date: 2025-06-12
Page: 1

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial endotoxins test	<0.5 endotoxins unit/ml	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**
Protein adsorption	Max 2.0 %	Complies
Cell adhesion	Max 1.0 %	Complies

**Only performed once

Sterility is obtained by ethylene oxide gas according to ISO 11135-1 "Sterilization of health care products - Ethylene oxide", with SAL 10E-6.

Roskilde Site



Mette Mikkelsen
Quality Control Coordinator

Please note that this is an electronically generated signature

Certificate reviewed by
Roskilde Site



Mette Holst Petsch
Sr. QC Technician

Please note that this is an electronically generated signature